

Over **60** years of consistent excellence

Servo Controlled Electronic Universal Testing Machine

SAVE CLR

EN

mm

Model : UTES



E-SERVO

Microprocessor Based Panel



Loading accuracy as high as $\pm 1\%$



Motor driven threaded columns for quick & effortless adjustment of lower cross-head-to facilitate rapid fixing of test specimen



Best in Class Analysis Software





Features :

- Loading accuracy as high as ± 1%
 Servo controlled Motorized Valve incorporating of control modes -Standard manual control, Potentiometric control, Load rate control, Elongation rate control, Load hold mode, Auto start & initial valve open start.
- Straining at variable speeds to suit a wide range of materials. Printer & PC graphs enable study the behavior of the material.
- Motor driven threaded columns for quick effortless adjustment of lower cross-head-to facilitate rapid fixing of test specimen.
- Simplicity in reading because of digital readouts.
- Wide range of standard and special accessories.
- Easy change from plain to threaded and screwed specimens.
- Large effective clearance between columns enables testing of standards specimens as well as structures.
- Simple controls for ease of operation.
- Robust straining frame of an extremely rigid construction.
- Safe operation ensured by means of safety devices.
- Fully enclosed and protected pressure transducer.
- RS 232 serial port for PC control of Machine.
- Load Capacity : 100 kN, 200 kN, 400 kN, 600 kN & 1000 kN.

Application:

'FIE' Electronic Universal Testing Machine is designed for testing Ferrous & Non-Ferrous materials under tension, compression bending, transverse and shear loads. Hardness test on metals can also conducted.

Machine Consists of -

Straining unit :

This consists of a cylinder motor with chain and sprocket drive and a table coupled with the ram of the hydraulic cylinder, mounted on to a robust base.

The cylinder and the ram are individually lapped to eliminate friction. The upper cross-head is rigidly fixed to the table by two strengthened columns.

The lower cross-head is connected to two screwed columns which are driven by a motor. Axial loading of the ram is ensured by relieving the cylinder and ram of any possible side loading by the provision of ball seating.

A Rotary Encoder of minimum resolution of 0.01mm is provided to measure the deformation of the specimen.

Tension test is conducted by gripping the test specimen between the upper and lower cross-heads.

Compression, transverse, bending, shear and hardness tests are conducted between the lower cross-head and the table.

The lower cross-head can be raised or lowered rapidly by operating the screwed columns, thus facilitating ease of fixing of the test specimen.

Control Panel :

The Control Panel consists of a power pack complete with drive motor and an oil tank, control valves and electronic display unit.

Power Pack :

The power pack generates the maximum pressure of 200 kgf/cm2. The hydraulic pump provides continuously non-pulsating oil flow. Hence the load application is very smooth.

Hydraulic Controls :

Hand operated release valve & motorized control valve are used to control the flow to and from the hydraulic cylinder. The regulation of the oil flow is infinitely variable. Incorporated in the hydraulic system is a regulating valve, which maintains a practically constant rate of piston movement.

Control by this valve allows extensometer reading to be taken.

Principle of Operation for -

Model : UTES

UTM right control valve is Servo Controlled in close loop mode as per mode selection. Following control modes available :

- 1) Standard Manual control
- 2) Potentiometric control
- 3) Load rate control
- 4) Elongation rate control
- 5) Load hold mode

Also Auto start, Potentometric start & initial value open start options are available for test start to take care of slippage & different specimen types. Load is applied by a hydrostatically lubricated ram.

Main cylinder pressure is transmitted to the pressure transducer housed in the control panel.

The transducer gives the signal to the electronic display unit, corresponding to the load exerted by the main ram.

Simultaneously the encoder fitted on the straining unit gives the mechanical displacement.

Electronic Control Panel (FIE SERVO) :



- In electronic panel 8085 microprocessor is used for basic UTM operation & an 8 bit dedicated micro controller is used for close loop & Servo control.
- Panel is having parallel printer port.
- RS 232 C serial port for PC interface.
- 2 lines x 8 digits 7 segment display & membrane keyboard for data entry.
- 20 data sets and 50 results storage.

Accuracy and Calibration :

FIE Electronic Universal testing machine is closely controlled for sensitivity, accuracy and calibration during every stage of manufacture.

Machine is calibrated over each of its measuring range in accordance with the procedure laid down in British standards 1610: Part1: 1992 and IS 1828: Part1: 1991.

FIE Electronic Universal Testing Machine complies with Grade "A" of BS: 1610:Part1:1992 and class 1 of IS-1828-Part1:1991.

Technical Specifications for -Servo Controlled Electronic Universal Testing Machines Model – UTES

MODEL	UNIT	UTES 10	UTES 20	UTES 40	UTES 60	UTE 100	UTES 120	UTES 150	UTES 200
Maximum Capacity	kN	100	200	400	600	1000	1200	1500	2000
Measuring range	kN	0-100	0-200	0-400	0-600	0-1000	0 - 1200	0 - 1500	0-2000
Load resolution (20000 counts full scale)	N	5	10	20	30	50	60	75	100
Load range with accuracy of Measurement +/-1%	kN	2 to 100	4 to 200	8 to 400	12 to 600	20 to 1000	24-1200	30-1500	40-2000
Resolution of piston movement (Displacement)	mm	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
Clearance for tensile test (At fully descended working piston)	mm	50-700	50-700	50-700	50-800	50-850	50-850	50-850	50-850
Clearance for compression test (At fully descended working piston)	mm	0-700	0-700	0-700	0-800	0-850	0-850	0-850	0-850
Clearance between columns	mm	500	500	500	600	750	750	750	750
Ram Stroke	mm	150	200	200	250	250	250	250	250
Straining/ Piston Speed (at no load)	mm/min	0-300	0-150	0-150	0-100	0-80	0-65	0-50	0-45
CONNECTED LOAD									
Power	HP	1.33	1.33	2.33	2.5	3.5	3.5	5	6
V		400-440	400-440	400-440	400-440	400-440	400-440	400-440	400-440
Φ		3	3	3	3	3	3	3	3
STANDARD ACCESSORIES									
FOR TENSION TEST									
Clamping jaws for round specimens		10-20	10-20	10-25	10-25	10-25	10-25	10-25	10-25
of diameter	mm	20-30	20-30	25-40	25-40	25-45	25-45	25-45	25-45
Clamping jaws for flat specimens	mm	0-10	0-10	0-15	0-15	0-22	0-20	0-20	0-20
of thickness		10-20	10-20	15-30	15-30	22-44	20-40	20-40	20-40
Width	mm	50	50	65	70	70	70	70	70
FOR COMPRESSION TEST									
Pair of Compression Plates of Dia.	mm	120	120	120	120	160	160	160	160
FOR TRANSVERSE TEST									
Table with adjustable rollers									
Width of rollers	mm	160	160	160	160	160	160	200	200
Diameter of Rollers	mm	30	30	30	50	50	50	50	50
Maximum clearance between supports	mm	500	500	500	600	800	800	800	800
Radius of punch tops	mm	6,12	6,12	12,16	16,22	16,22	16, 22	16, 22	16, 22

• Due to constant R& D specifications & features are subject to change without notice.

• Colour scheme subject to confirm at the time of order.




Software Package -

'FIE' Tension & Compression test Software package for Model : UTES

Input Data

ret made select.	Plupture (% Presk)	50	
Manual Control	Pretonal (% Full Scale)	D T	
Patentionetric Control	Sale Louit (6N)	408	
and flate Cornal	Hold time (Sec.)	18	
Impetion Nate control	Local State (6N/men)	200	
ned Hold	Stress flate (M/agam/eis) 20		
roor Bate Control	Emogetion Rate (mm/mm)	2	
tran Rate Castrol	Strain State (Scateaus/man)	0.5	
	Initial Velve Open . 46-86%	25	
t forst on do Slavt 6. etanl dial Valve Opon	Servo Control Modes Data Entry	•	

Wide range of Control Modes with Input Data entry.




Real time values of load & displacement/extension used for plotting its graph & can be seen real time during testing.

Graph Superimpose with Point Tracing

Construction C	a lent		1	Foler unit dat		office willing
Bar (1) Control Control State 200 200 200				-		Debit Linux On Groph Departmipsion antioci
Howards 1200 14 1200 Ling 1400 1200 14 1200 Ling 1400 1200 1200 Name 1200 14 1200			Grigh	Load Vo Coleman	1	
All and 10 2.01 4.01 COL 31400 1.01 1.00 All Annual Line 2.011 1.00 Max Line 1.02 1.00 Max Line 1.011 1.01	Contractory of	1010	3			
List 21100 List 2015 M 100 List 2016 M 100 100 List 2016 M 100 100 Million 100 M 100	Result					
Association 1711.54 100 Mail Statem 1.30 rpc 1.00 Mail Table of the statem 1.00 1.00 Mail Statem 1.01 statem 1.01 Mail Statem 1.01 statem 1.01 Mail Statem 1.01 statem 1.01	Contraction of the local division of the loc					
Abilitation 1.00 mm 1.00 mm Abilitation 1.00 mm 0.00 mm	and the second					
Image: 17 Dis agram 1.000 M Toma: 4.201 Distribution 1.000 M Toma: 1.001 Distribution 1.000 M Toma: 1.001 Distribution 4.000 M Toma: 1.001 Distribution 1.001				10000		
Image C.D. Status 1.26 Status D.L.S. Status 1.26 Status T.D. Status 6.00	No.					
All Constant	In law			1/		
Million 110 100 100 113 Million 110 100 100 113 Million 110 100 100 113 Million 110 100 100	This same	300	100	1.1		
All The Control of Con		NO.				
13 Martines 2 J 20 Strates 23 Martines 2 St 24 Name 24 Martines 2 St 24 Name 25 Martines 2 St 25 Name 25 Martines 2 St 26 Name 25 Martines 2 St	al sec	THEFT		1		
Anna hap and a second s		31/21/4	10.000			
		6854	AL			
Baserian 198 198 198 199 199 199 199 199 199 199	And in street				15 K.	1 K
	Butter 14	100		1.000	1000 - 1000	1000-1000-1000-1000-

Graph comparison, superimpose of batch sample tests with point tracing facility.

Load Vs Extension



Displays graph for extensometer readings against load with proof lines for proof stress.

Statistics



Batch test reports with statistics for batch sample results.

Windows based software available for -Online graph on PC, Data analysis, Statistics, Point tracing, superimposing graphs to compare with standard, zooming graph etc.

Manufactured By : **Fuel Instruments & Engineers Pvt. Ltd.** Sold And Serviced by : SUZUKI INSTRUMENTS **Cell : +91 9594084085, E-mail : sales@simaterialtestingservices.com**

Web : www.simaterialtestingservices.com